

Certificate of Analysis
for
CDS SDB-XC 1ml / 4mm

CDS Product ID: 98-0604-0201-7

Certificate Date: 05-18-21

CDS Lot #'s: 750063D

Property (Units)	Avg thick (mil)	Avg flow (min/L)	Recovery
			(%)
Test Method	5136	5133	5097
Upper Specification	22	10	
Lower Specification	18	3	80
Membrane Lot #	Data		
750063D	21	4	97

This is to certify that the material shipped on this order is in conformance with CDS specifications.

CDS Company Quality Assurance

 TA

Important Notice to Purchaser:

The seller's and manufacturer's only obligation will be to replace the quantity of the product proved to be defective. Neither the seller nor CDS will be liable for any injury, loss or damage, direct or consequential, arising out of the use of or the inability to use the product. Before using, the user must determine the suitability of the product for his or her intended use.