

Certificate of Analysis

for

CDS Chelator SPE 47mm Disks

CDS Product ID: 98-0604-0238-9

Certificate Date: 05/23/22

CDS Lot #' 530005C

SAP ID: 7100029525

Property (Units)	Avg thick (mil)	Avg flow (min/L)	Recovery (%)
Test Method	5136	5133	5150
Upper Specification	22	16	
Lower Specification	18	8	80
Membrane Lot #	Data		
530005C	23	11.5	100

This is to certify that the material shipped on this order is in conformance with 3M specifications.

3M Company Quality Assurance  TA

Important Notice to Purchaser:

The seller's and manufacturer's only obligation will be to replace the quantity of the product proved to be defective. Neither the seller nor 3M will be liable for any injury, loss or damage, direct or consequential, arising out of the use of or the inability to use the product. Before using, the user must determine the suitability of the product for his or her intended use.