

Certificate of Analysis

for

RADIUM 47MM DISK

CDS Product ID: 98-0405-0065-8EA

Certificate Date: 25-Oct-18

CDS Lot #'s: 440029D

SAP ID: 7000059475

Property (Units)	Avg thick (mil)	Avg flow (min/L)	Recovery
			(%)
Specification	>22	4-10	<u>>95</u>
Membrane Lot #	Data		
440029D	21.8	7.8	100

This is to certify that the material shipped on this order, with the preceeding analysis, is in conformance with CDS specifications.

CDS Company Quality Assurance



TA

Important Notice to Purchaser:

The seller's and manufacturer's only obligation will be to replace the quantity of the product proved to be defective. Neither the seller nor CDS will be liable for any injury, loss or damage, direct or consequential, arising out of the use of or the inability to use the product. Before using, the user must determine the suitability of the product for his or her intended use.