

Certificate of Analysis

for

CDS Chelator 6ml / 10mm

CDS Product ID: 98-0604-0701-6

Certificate Date:08/12/2024

CDS Lot #'s: 550002C

Property (Units)	Avg thick (mil)	Avg flow (min/L)	Recovery
			(%)
Test Method	5136	5133	5150
Upper Specification	22	16	
Lower Specification	18	8	80
Membrane Lot #	Data		
550002C	22	15	96

This is to certify that the material shipped on this order is in conformance with CDS specifications.

CDS Company Quality Assurance



TA

Important Notice to Purchaser:

The seller's and manufacturer's only obligation will be to replace the quantity of the product proved to be defective. Neither the seller nor CDS will be liable for any injury, loss or damage, direct or consequential, arising out of the use of or the inability to use the product. Before using, the user must determine the suitability of the product for his or her intended use.