

Certificate of Analysis

for

3M MPC 96 Well Plate

3M Product ID: 70-2007-3983-0EA

Certificate Date: 24-Jun-19

3M Lot #'s: 690119D

Property (Units)	Avg thick (mil)	Avg flow (min/L)	Recovery
			(%)
Test Method	5136	5133	5144
Upper Specification	32	4	
Lower Specification	28	1.5	80
Membrane Lot #	Data		
690119D	31.2	3.6	100.0

This is to certify that the material shipped on this order is in conformance with 3M specifications.

3M Company Quality Assurance



TA

Important Notice to Purchaser:

The seller's and manufacturer's only obligation will be to replace the quantity of the product proved to be defective. Neither the seller nor 3M will be liable for any injury, loss or damage, direct or consequential, arising out of the use of or the inability to use the product.

Before using, the user must determine the suitability of the product for his or her intended use.